

AD0641571

-A

2886

2886

# A SIMPLE PORTABLE SONOMICROMETER

MERRILL B. KARDON  
HUGH F. STEGALL, Captain, USAF, MC  
HUBERT L. STONE, Ph.D.

20090501 024

November 1966

USAF School of Aerospace Medicine  
Aerospace Medical Division (AFSC)  
Brooks Air Force Base, Texas

Qualified requesters may obtain copies of this report from DDC. Orders will be expedited if placed through the librarian or other person designated to request documents from DDC.

When U. S. Government drawings, specifications, or other data are used for any purpose other than a definitely related government procurement operation, the government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Distribution of this document is unlimited.

**A SIMPLE PORTABLE SONOMICROMETER**

**MERRILL B. KARDON**

**HUGH F. STEGALL, Captain, USAF, MC**

**HUBERT L. STONE, Ph.D.**

## FOREWORD

The sonomicrometer described in this report was constructed to fulfill a need for continuous registration of cardiac dimensions in conscious dogs exposed to acceleration stress. The instrument was developed in the Biodynamics Branch, USAF School of Aerospace Medicine, between March 1966 and July 1966. The work was accomplished under task No. 793003 and was supported in part by National Aeronautics and Space Agency contract No. T-37761-G. The paper was submitted for publication on 24 August 1966.

All experiments were conducted according to the "Principles of Laboratory Animal Care" of the National Society for Medical Research.

The authors extend their appreciation to Dr. V. S. Bishop, Roy Turner, and Master Sergeant Ben Wiggins for their assistance and suggestions.

This report has been reviewed and is approved.

A handwritten signature in cursive script, reading "James B Nuttall". The signature is written in dark ink and is positioned above the printed name and title.

JAMES B. NUTTALL  
Colonel, USAF, MC  
Commander

## ABSTRACT

The design and construction of a relatively simple and inexpensive ultrasonic dimension gage (sonomicrometer) are described in detail. Two piezoelectric crystals are installed surgically across the organ to be studied, and the transit time for a burst of ultrasound to pass from one crystal to the other is measured. The device is easily calibrated when sound velocity in the medium is known. By use of integrated circuits throughout, it may be duplicated for less than \$55. Possible applications other than continuous measurements of organ size are discussed.



# A SIMPLE PORTABLE SONOMICROMETER

## I. INTRODUCTION

Continuous, dynamic, direct measurements of organ dimensions in man and animals are necessary in order to understand changes in organic function. In particular, reactions of the heart to various stresses can be detected by measurements of ventricular diameter (1).

Among the technics that have been suggested as means of obtaining organ dimensions in the conscious animal, Peterson (2) lists five:

1. Resistance strain gages.
2. Inductance strain gages.
3. Impedance plethysmographs.
4. Optical and roentgenographic technics.
5. Sonic (i.e., ultrasonic) methods.

The latter offer several advantages. They are inherently calibrated since transit time is directly proportional to distance when velocity is known. Drift is entirely an electronic problem because the sensors do not change properties. The transducer is light and the two halves are not physically linked, thus allowing free motion of the organ under measurement.

Instead of using the arrival time of an ultrasonic echo, as is done in conventional sonar, Baker (3) suggested that transit time of sound between two piezoelectric crystals mounted on either side of an organ could be used to indicate changes in the distance between the crystals. Using this principle, we developed a "sonomicrometer," which eliminates the major objection to using sonic technics—i.e., the need for complex, cumbersome circuitry to process the information. In order to obtain either telemetered or hard-wire-transmitted measurements from an animal under stress, it was also desirable to have a lightweight, battery-operated instrument.

## II. DESIGN

Figure 1 is a block diagram of the designed instrument and shows the technic of information processing used. A pulse from an avalanche mode oscillator ("pinger") excites the transmitting crystal and also sets a monostable multivibrator. The trailing edge of the monostable sets a bistable multivibrator about 3 to 4  $\mu$ sec. after the initial pulse. Ultrasonic energy in the form of a burst of 5 mHz sound waves passes through the organ and is detected by the receiving crystal. The received signal is then used to reset the bistable multivibrator, completing one timing cycle. The bistable output is filtered through a conventional R-C network for recording.

Tracing one complete timing cycle yields a clearer picture of the function of each element. Each cycle begins with the ultrafast pulse (less than  $10^{-8}$  seconds rise time) from the blocking oscillator. Ringing appears on this pulse because it drives the resonant piezoelectric element. Although minimized by the very low output impedance of the pulse generator, some ringing is capacitatively coupled and appears immediately at the receiver. If the bistable multivibrator were "set" by the original oscillator pulse, it would be quickly "reset" by this capacitatively coupled signal. Baker (3) suggested that the receiver be gated "off" during this early period in order to avoid the problem, but a simpler solution was found by delaying the "set" pulse by means of a monostable multivibrator whose width was greater than this capacitatively coupled pulse. Moreover, as will be seen below, monostable width could be varied for zero suppression when pulsatile changes in distance were to be recorded.




FIGURE 1

Block diagram with waveforms illustrating the output of each stage: (1) shows the "pinger" pulse, which drives the transmitting crystal; (2) is the monostable multivibrator output (trailing edge "sets" the bistable); (3) is the receiver output, showing "pinger" pulse feedthrough and received packet, which "resets" the bistable; (4) is the bistable multivibrator output ( $t_2$  varies with transit time).

Low-frequency components at approximately 0.5 mHz, apparently the result of oscillations other than in the thickness mode, were eliminated by a high-pass L-C filter between the receiver and the Schmitt trigger. The trigger detects the time when the receiver output rises above electronic noise in the system, forming a positive pulse which resets the bistable multivibrator. Ordinarily, this is adjusted to occur during the leading edge of the first or second positive half-cycle, whichever is most stable. With this, the bistable multivibrator is reset, and thus a timing cycle is completed.

Bistable output is a rectangular pulse whose duration is equal to the acoustic delay between transmitted and received pulse less the electronic delay of the monostable multivibrator. Since pulse repetition rate is constant, the distance information appears as duty cycle modulation, and may be recovered by a simple R-C filter. A repetition rate of 5,000 measurements per second was selected in order to maintain good sensitivity and reduce crosstalk with other

ultrasonic sensors. Increasing the monostable delay reduces the pulse width by a fixed amount, thus allowing easier inspection of pulsatile changes in organ size.

Figure 2 (A, B, C, D) is the schematic for the finished design. With the exception of the transmitter ("pinger"), all stages use integrated circuits available at a cost of \$4 to \$15 each. Cost for all parts for the entire circuit is less than \$55, and may soon be even lower as the cost of integrated circuits continues to fall. A competent technician should be able to duplicate this instrument within a relatively brief time since no special construction techniques appear necessary other than the usual precautions in high-frequency design: short leads with careful lead dress, decoupling between stages, etc.

### III. CONSTRUCTION

As pictured in figure 3, the sonomicrometer was constructed in separate stages with each stage on a separate copper-clad board. The






FIGURE 2A

*Schematic of complete sonomicrometer. All resistors are  $\frac{1}{4}$  watt; all capacitors are listed in microfarads unless otherwise stated.*

copper coating was left intact except around the isolated interconnection pins. All grounds were carried directly to the intact copper coating. Tied together by two  $\frac{1}{8}$ -inch brass rods with threaded ends, the assembly was placed inside a brass box (fig. 4). Connections were included for attaching crystals, supplying power, and testing the output from each stage. A lid (not shown) completed the construction.

Piezoelectric transducers were constructed by use of a ceramic material resonant at approximately 5 MHz in its thickness mode. Cut to 2 to 5 mm. round, the crystals were placed in specially cast holders as shown in figure 5. The crystal holders were cast from small, hollow, nontoxic polystyrene beads in a brass mold heated to 150° C. for 10 minutes; they have about the same consistency as hot-drink

cups. A hole drilled from one end through the holder toward its center allowed for the passage of a pair of wires to connect to the crystal faces. Several additional holes were drilled around its circumference for suture placement. A spherical ( $\frac{1}{8}$ -inch to  $\frac{1}{4}$ -inch radius of curvature) polyester lens was glued to the outer face of each crystal to reduce directivity and broaden the ultrasonic beam.

Wires were soldered to the crystal faces with a low temperature solder, melting at approximately 100° C., and a soldering iron just warm enough to melt it. First, a small section of the surface coating was scraped from the crystal, thus exposing the metal plating underneath. A tiny drop of special flux supplied with the solder was applied, and a very small dot of solder was then carried on the tip of the






FIGURE 2B

*Internal components and lead connections for Fairchild  $\mu A 702 C$  integrated circuit used as radio frequency amplifier.*

iron to the prepared spot. A quick touch deposited the solder dot on the prepared surface, and a pre-tinned wire was laid over this spot and, in turn, fused to it with the iron. With some practice, the operator learned to apply a relatively small amount of heat to the crystal, thereby reducing the chance of thermal damage to the piezoelectric material.

#### Instrument characteristics

A jig was constructed to hold two crystal transducers at distances which could be varied with a micrometer. The entire assembly was placed under water at room temperature, and output of the sonomicrometer was plotted as a function of distance between the crystals. Figure 6 illustrates that output varied linearly

with crystal spacing, and that overall sensitivity was approximately 6.7 mv./mm. change in spacing. After batteries had aged slightly, output drift was less than 3 mv. in 45 minutes; this appeared to be due entirely to a gradual reduction of supply voltage from the batteries. Absolute accuracy, using published figures for sound velocity in water, was better than 5%.

No driver was available for adequate determination of overall frequency response; however, it would appear reasonable that at this sampling rate (5,000 measurements per second), frequency response would depend entirely on output filter characteristics. The low-pass R-C filter used has a corner frequency of 16 Hz, and should pass unimpeded biologic information, which may be assumed to have a




FIGURE 2C  
Circuit diagram for Fairchild  $\mu L914$ .




FIGURE 2D  
Basing diagrams for  $\mu L914$  and  $\mu A702C$ .




FIGURE 3

*Each stage is built on a separate board. Connections are made from the integrated circuit socket to isolated pins. Copper coating used for ground is intact except around pins.*




FIGURE 4

*Completed sonomicrometer shows stacking of circuit cards using plastic separators. Pin jacks on lower edge of package are for test and signal monitoring purposes. One transducer pair is shown at left.*






FIGURE 5

*Transducer construction showing the crystal holder and diffusing lens. The holder, cast of polystyrene beads, also acts as acoustical backing.*




FIGURE 6

*Sonomicrometer output as a function of crystal spacing; sensitivity is approximately 6.5 mv./mm. change in spacing.*

maximum appreciable harmonic content of one-third this value. Phase shift may be assumed negligible by a similar argument.

### Calibration

The velocity of ultrasound in biologic structures is approximately that for water at

37° C.—i.e., about  $1.55 \times 10^6$  mm./sec. A calibration fixture was constructed of a rectangular  $\frac{1}{4}$ -inch Lucite sheet whose ends had been milled in steps (fig. 7). Crystals were glued to the Lucite faces, and transit time between each pair of crystals was measured accurately on an oscilloscope. Though the velocity of sound in Lucite is almost twice that






FIGURE 7

*Calibration fixture with known water equivalent distances in millimeters. The fixture is made of Lucite and silicone rubber.*

in water, a "water equivalent" was readily calculated and each crystal pair so labeled. During recordings from animal structures, these crystals with a known ultrasonic delay could be inserted at the instrument input and its output displayed on the recorder. Two such points were enough to calibrate each record. The calibrator was constructed to yield equivalent distances in water from 40 to 60 mm. in 5-mm. increments. This distance was sufficient to measure most of the canine hearts examined.

#### IV. DISCUSSION

##### Application

This instrument has been restricted thus far to making continuous measurements of left

ventricular diameter in conscious dogs. Figure 8 shows records obtained from one dog at rest; pulmonary blood flow (by means of an electromagnetic flowmeter) and electrocardiogram were also recorded.

Crystal transducers were placed surgically across the left ventricle in healthy mongrel dogs. The heart was exposed through a left anterior oblique approach; the pericardium was incised; and the heart was lifted anteriorly while one crystal was sutured to the posterior surface of the left ventricle. The best position for the anterior crystal was determined by exploring the ventricular surface until the largest, most consistent ultrasonic burst was detected, and then the crystal was sutured into place. Both crystals were about 1 to 2 cm.




FIGURE 8

*Simultaneous recordings of ECG, pulmonary artery flow (by means of an implanted electromagnetic flow-meter probe), and left ventricular diameter in a conscious dog. Paper speed, 25 mm./sec.*

below the atrioventricular groove, and were directed across the greatest diameter of the left ventricle. Within the first day or two after surgery, the received signal appeared to reach its maximum. While some animals have thus far yielded usable records for several weeks following implantation, prediction of anticipated useful life cannot be made until more animals are studied. No adverse reaction to the transducers themselves has been seen.

Similar instruments have been used by others to measure dimensions across liver and spleen (4) or to estimate blood volume changes in a portion of mesentery (5). Crystals across a man's arm or leg have been used to measure diameter continuously, as a possible substitute for conventional plethysmography (6), and other possible applications are limited only by the imagination of the investigator.

#### Disadvantages

The measurement of a linear dimension is not by any means a substitute for measurement of an organ's volume although it may indirectly

indicate changes in volume. Moreover, crystals applied to the outside of the dog heart measure thickness of both the myocardium and the contained blood. Usually the latter is the variable of interest.

Crystal design and application are not yet optimized. The leading edge of the received ultrasonic burst must rise out of the background noise fast enough to trigger reliably; this demands a reasonable degree of acoustical coupling between transmitting and receiving crystals, coupling which may be readily achieved on the operating table, but might gradually fade over the next few days as the size and shape of the dog heart return to normal. Other organs less affected by the surgical procedure itself should present fewer problems.

Some degree of skill in maintaining the instrument is still necessary. An oscilloscope with adequate frequency response (to 5 mHz) is essential. The waveforms at receiver, Schmitt, and bistable outputs should be checked repeatedly to assure that the triggering point



has not fluctuated. A multichannel oscilloscope is desirable when adjusting monostable delay and trigger points.

#### Future development

The device described here is sufficiently small to allow telemetered recordings from an animal roaming freely about the laboratory. This will be attempted when a more compact construction is complete. Further investigation of crystal transducer construction has

already begun. Some other lens materials and designs have been examined. A careful study of ultrasonic dispersion by such technics seems indicated.

Other organs than the heart will be examined as well. Pulsations of kidney, liver, and mesenteric beds as perfusing pressure is altered can furnish valuable information about their function. Visceral motility in conscious animals could be easily studied by this technic.

#### REFERENCES

1. Rushmer, R. F. Cardiovascular dynamics. Philadelphia, Pa.: W. B. Saunders, 1961.
2. Peterson, L. H. General types of dimensional transducers. *In* Rushmer, R. F. (ed.). *Methods of medical research*, pp. 5-23. Chicago, Ill.: Year Book Medical Publishers, Inc. 1966.
3. Baker, D. W. Sonocardiometer. *In* Rushmer, R. F. (ed.). *Methods of medical research*, pp. 31-34. Chicago, Ill.: Year Book Medical Publishers, Inc., 1966.
4. Guntheroth, W. G., and G. L. Mullins. Liver and spleen as venous reservoirs. *Amer. J. Physiol.* 204:35 (1963).
5. Mullins, G. L., and W. G. Guntheroth. Continuous recording of changes in mesenteric blood volume. *Ultrasonics* 4:42 (1966).
6. Stegall, H. F., J. M. Ofstad, V. E. Simmons, and R. F. Rushmer. An integrated facility for peripheral vascular studies in man. (Abstract) *Proc. 17th Ann. Conf. Eng. Med. Biol.* 6:75 (1964).

## DOCUMENT CONTROL DATA - R&amp;D

(Security classification of title, body of abstract and indexing annotation must be entered when the overall report is classified)

1. ORIGINATING ACTIVITY (Corporate author) USAF School of Aerospace Medicine Aerospace Medical Division (AFSC) Brooks Air Force Base, Texas		2a. REPORT SECURITY CLASSIFICATION Unclassified	
		2b. GROUP	
3. REPORT TITLE  A SIMPLE PORTABLE SONOMICROMETER			
4. DESCRIPTIVE NOTES (Type of report and inclusive dates) Mar. - July 1966			
5. AUTHOR(S) (Last name, first name, initial) Kardon, Merrill B. Stegall, Hugh F., Captain, USAF, MC Stone, Hubert L.			
6. REPORT DATE Nov. 1966		7a. TOTAL NO. OF PAGES 10	7b. NO. OF REFS 6
8a. CONTRACT OR GRANT NO NASA Contract No. T-37761-G		9a. ORIGINATOR'S REPORT NUMBER(S)  SAM-TR-66-96	
b. PROJECT NO.			
c. Task No. 793003		9b. OTHER REPORT NO(S) (Any other numbers that may be assigned this report)	
d.			
10. AVAILABILITY/LIMITATION NOTICES  Distribution of this document is unlimited.			
11. SUPPLEMENTARY NOTES		12. SPONSORING MILITARY ACTIVITY USAF School of Aerospace Medicine Aerospace Medical Division (AFSC) Brooks Air Force Base, Texas	
13. ABSTRACT  The design and construction of a relatively simple and inexpensive ultrasonic dimension gage (sonomicrometer) are described in detail. Two piezoelectric crystals are installed surgically across the organ to be studied, and the transit time for a burst of ultrasound to pass from one crystal to the other is measured. The device is easily calibrated when sound velocity in the medium is known. By use of integrated circuits throughout, it may be duplicated for less than \$55. Possible applications other than continuous measurements of organ size are discussed.			



14. KEY WORDS	LINK A		LINK B		LINK C	
	ROLE	WT	ROLE	WT	ROLE	WT
Physiology						
Bioinstrumentation						
Ultrasound, diagnostic						
Sonocardiometer						
Cardiac volume						
Cardiac diameter						

## INSTRUCTIONS

1. **ORIGINATING ACTIVITY:** Enter the name and address of the contractor, subcontractor, grantee, Department of Defense activity or other organization (*corporate author*) issuing the report.

2a. **REPORT SECURITY CLASSIFICATION:** Enter the overall security classification of the report. Indicate whether "Restricted Data" is included. Marking is to be in accordance with appropriate security regulations.

2b. **GROUP:** Automatic downgrading is specified in DoD Directive 5200.10 and Armed Forces Industrial Manual. Enter the group number. Also, when applicable, show that optional markings have been used for Group 3 and Group 4 as authorized.

3. **REPORT TITLE:** Enter the complete report title in all capital letters. Titles in all cases should be unclassified. If a meaningful title cannot be selected without classification, show title classification in all capitals in parenthesis immediately following the title.

4. **DESCRIPTIVE NOTES:** If appropriate, enter the type of report, e.g., interim, progress, summary, annual, or final. Give the inclusive dates when a specific reporting period is covered.

5. **AUTHOR(S):** Enter the name(s) of author(s) as shown on or in the report. Enter last name, first name, middle initial. If military, show rank and branch of service. The name of the principal author is an absolute minimum requirement.

6. **REPORT DATE:** Enter the date of the report as day, month, year; or month, year. If more than one date appears on the report, use date of publication.

7a. **TOTAL NUMBER OF PAGES:** The total page count should follow normal pagination procedures, i.e., enter the number of pages containing information.

7b. **NUMBER OF REFERENCES:** Enter the total number of references cited in the report.

8a. **CONTRACT OR GRANT NUMBER:** If appropriate, enter the applicable number of the contract or grant under which the report was written.

8b, 8c, & 8d. **PROJECT NUMBER:** Enter the appropriate military department identification, such as project number, subproject number, system numbers, task number, etc.

9a. **ORIGINATOR'S REPORT NUMBER(S):** Enter the official report number by which the document will be identified and controlled by the originating activity. This number must be unique to this report.

9b. **OTHER REPORT NUMBER(S):** If the report has been assigned any other report numbers (*either by the originator or by the sponsor*), also enter this number(s).

10. **AVAILABILITY/LIMITATION NOTICES:** Enter any limitations on further dissemination of the report, other than those

imposed by security classification, using standard statements such as:

- (1) "Qualified requesters may obtain copies of this report from DDC."
- (2) "Foreign announcement and dissemination of this report by DDC is not authorized."
- (3) "U. S. Government agencies may obtain copies of this report directly from DDC. Other qualified DDC users shall request through \_\_\_\_\_."
- (4) "U. S. military agencies may obtain copies of this report directly from DDC. Other qualified users shall request through \_\_\_\_\_."
- (5) "All distribution of this report is controlled. Qualified DDC users shall request through \_\_\_\_\_."

If the report has been furnished to the Office of Technical Services, Department of Commerce, for sale to the public, indicate this fact and enter the price, if known.

11. **SUPPLEMENTARY NOTES:** Use for additional explanatory notes.

12. **SPONSORING MILITARY ACTIVITY:** Enter the name of the departmental project office or laboratory sponsoring (*paying for*) the research and development. Include address.

13. **ABSTRACT:** Enter an abstract giving a brief and factual summary of the document indicative of the report, even though it may also appear elsewhere in the body of the technical report. If additional space is required, a continuation sheet shall be attached.

It is highly desirable that the abstract of classified reports be unclassified. Each paragraph of the abstract shall end with an indication of the military security classification of the information in the paragraph, represented as (TS), (S), (C), or (U).

There is no limitation on the length of the abstract. However, the suggested length is from 150 to 225 words.

14. **KEY WORDS:** Key words are technically meaningful terms or short phrases that characterize a report and may be used as index entries for cataloging the report. Key words must be selected so that no security classification is required. Identifiers, such as equipment model designation, trade name, military project code name, geographic location, may be used as key words but will be followed by an indication of technical context. The assignment of links, rules, and weights is optional.

USAF SCHOOL OF AEROSPACE MEDICINE  
AEROSPACE MEDICAL DIVISION (AFSC)  
BROOKS AFB TEX 78235

UNITED STATES AIR FORCE  
OFFICIAL BUSINESS

33246  
6570 AMRL (MRHE/DR. J.M. CHRISTENSEN)  
WRIGHT PATTERSON AFB OHIO 45433